

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012725\
 Data File : VX044721.D
 Acq On : 27 Jan 2025 14:09
 Operator : JC/MD
 Sample : Q1173-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-192-GW-200-202

Quant Time: Jan 28 04:12:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 16 01:16:09 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	143949	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	289107	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	257515	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	112964	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	103939	47.836	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	95.680%		
35) Dibromofluoromethane	5.379	113	93332	50.816	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	101.640%		
50) Toluene-d8	8.647	98	348739	54.456	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	108.920%		
62) 4-Bromofluorobenzene	11.079	95	127861	53.558	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	107.120%		
Target Compounds						
16) Acetone	2.380	43	6091	7.085	ug/l	97
30) Chloroform	5.099	83	2310	0.677	ug/l	99
44) Trichloroethene	7.129	130	3350	1.759	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012725\
Data File : VX044721.D
Acq On : 27 Jan 2025 14:09
Operator : JC/MD
Sample : Q1173-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-192-GW-200-202

Quant Time: Jan 28 04:12:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

