

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012725\
 Data File : VX044724.D
 Acq On : 27 Jan 2025 15:18
 Operator : JC/MD
 Sample : PB166212ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB166212ZHE#02

Quant Time: Jan 28 04:13:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 16 01:16:09 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	149280	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	295118	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	249832	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	106375	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	107748	47.819	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	95.640%	
35) Dibromofluoromethane	5.379	113	94468	50.387	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	100.780%	
50) Toluene-d8	8.646	98	355410	54.367	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	108.740%	
62) 4-Bromofluorobenzene	11.079	95	123453	50.658	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	101.320%	
Target Compounds						
						Qvalue
16) Acetone	2.385	43	60152	67.474	ug/l	92
18) Methyl Acetate	2.708	43	3370	1.225	ug/l	99
20) Methylene Chloride	2.782	84	2168	1.044	ug/l	95
25) 2-Butanone	4.586	43	7167	5.407	ug/l	98
43) Isopropyl Acetate	6.336	43	152587	29.844	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012725\
Data File : VX044724.D
Acq On : 27 Jan 2025 15:18
Operator : JC/MD
Sample : PB166212ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB166212ZHE#02

Quant Time: Jan 28 04:13:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

