

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012725\
 Data File : VX044725.D
 Acq On : 27 Jan 2025 15:42
 Operator : JC/MD
 Sample : PB166212ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB166212ZHE#03

Quant Time: Jan 28 04:14:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 16 01:16:09 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	144171	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	284703	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	249994	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	108791	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	104715	48.119	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	96.240%		
35) Dibromofluoromethane	5.379	113	91211	50.430	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	100.860%		
50) Toluene-d8	8.647	98	345617	54.803	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	109.600%		
62) 4-Bromofluorobenzene	11.079	95	122315	52.027	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	104.060%		
Target Compounds						
16) Acetone	2.380	43	50898	59.117	ug/l	91
25) 2-Butanone	4.580	43	6671	5.211	ug/l	96
43) Isopropyl Acetate	6.336	43	146387	29.679	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012725\
Data File : VX044725.D
Acq On : 27 Jan 2025 15:42
Operator : JC/MD
Sample : PB166212ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB166212ZHE#03

Quant Time: Jan 28 04:14:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

