

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012725\
 Data File : VX044726.D
 Acq On : 27 Jan 2025 16:04
 Operator : JC/MD
 Sample : Q1169-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ARS20-0006

Quant Time: Jan 28 04:14:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 16 01:16:09 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	141001	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	286144	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	250593	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	106831	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	105047	49.357	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.720%
35) Dibromofluoromethane	5.379	113	91548	50.361	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.720%
50) Toluene-d8	8.647	98	345312	54.479	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	108.960%
62) 4-Bromofluorobenzene	11.079	95	119925	50.754	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.500%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	53631	63.691	ug/l	99
18) Methyl Acetate	2.703	43	2853	1.098	ug/l	96
25) 2-Butanone	4.580	43	7123	5.689	ug/l	94
43) Isopropyl Acetate	6.342	43	129831	26.190	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012725\
Data File : VX044726.D
Acq On : 27 Jan 2025 16:04
Operator : JC/MD
Sample : Q1169-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ARS20-0006

Quant Time: Jan 28 04:14:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

