

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012725\
 Data File : VX044727.D
 Acq On : 27 Jan 2025 16:28
 Operator : JC/MD
 Sample : Q1169-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 REGULATOR-BUILDING-GRAVEL

Quant Time: Jan 28 04:15:01 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 16 01:16:09 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	148587	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	298945	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	254468	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	109772	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	111164	49.565	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.120%
35) Dibromofluoromethane	5.379	113	87866	46.266	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.540%
50) Toluene-d8	8.647	98	360726	54.474	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	108.940%
62) 4-Bromofluorobenzene	11.079	95	124947	50.615	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.220%
Target Compounds						
					Qvalue	
16) Acetone	2.385	43	48841	55.042	ug/l	93
18) Methyl Acetate	2.709	43	2803	1.024	ug/l	95
43) Isopropyl Acetate	6.336	43	143791	27.764	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012725\
Data File : VX044727.D
Acq On : 27 Jan 2025 16:28
Operator : JC/MD
Sample : Q1169-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
REGULATOR-BUILDING-GRAVEL

Quant Time: Jan 28 04:15:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

