

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012725\
 Data File : VX044731.D
 Acq On : 27 Jan 2025 18:00
 Operator : JC/MD
 Sample : Q1180-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Storage-Blank-WATER-REF-5

Quant Time: Jan 28 04:16:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 16 01:16:09 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	138064	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	281466	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	252440	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	106684	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	102440	49.156	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.320%
35) Dibromofluoromethane	5.379	113	90352	50.529	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.060%
50) Toluene-d8	8.647	98	344222	55.210	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	110.420%
62) 4-Bromofluorobenzene	11.079	95	123053	52.943	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	105.880%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012725\
Data File : VX044731.D
Acq On : 27 Jan 2025 18:00
Operator : JC/MD
Sample : Q1180-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Storage-Blank-WATER-REF-5

Quant Time: Jan 28 04:16:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

