

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012725\
 Data File : VX044732.D
 Acq On : 27 Jan 2025 18:23
 Operator : JC/MD
 Sample : Q1180-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Storage-Blank-WATER-REF-4

Quant Time: Jan 28 04:17:02 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 16 01:16:09 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	139447	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	282344	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	252421	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	105754	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	103047	48.957	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	97.920%		
35) Dibromofluoromethane	5.379	113	89910	50.126	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	100.260%		
50) Toluene-d8	8.647	98	344260	55.044	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	110.080%		
62) 4-Bromofluorobenzene	11.079	95	121898	52.283	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	104.560%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012725\
Data File : VX044732.D
Acq On : 27 Jan 2025 18:23
Operator : JC/MD
Sample : Q1180-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Storage-Blank-WATER-REF-4

Quant Time: Jan 28 04:17:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

