

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX012819\
 Data File : VX007269.D
 Acq On : 28 Jan 2019 18:58
 Operator : JC/SP
 Sample : K1011-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OR-01-012519-A

Quant Time: Jan 29 01:24:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	539282	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	826465	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	792037	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	387742	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	290565	51.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.12%	
35) Dibromofluoromethane	5.50	113	260116	49.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.30%	
50) Toluene-d8	8.71	98	1020985	52.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.40%	
62) 4-Bromofluorobenzene	11.13	95	359521	53.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.46%	
Target Compounds						
16) Acetone	2.45	43	64146	23.022	ug/l	99
20) Methylene Chloride	2.86	84	483269	91.664	ug/l	97
25) 2-Butanone	4.70	43	9656	2.422	ug/l	95
43) Isopropyl Acetate	6.46	43	18844	1.551	ug/l #	95
95) Naphthalene	13.83	128	35953	1.287	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX012819\
Data File : VX007269.D
Acq On : 28 Jan 2019 18:58
Operator : JC/SP
Sample : K1011-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OR-01-012519-A

Quant Time: Jan 29 01:24:59 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

