

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX012819\
 Data File : VX007271.D
 Acq On : 28 Jan 2019 19:52
 Operator : JC/SP
 Sample : K1254-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SOIL-1

Quant Time: Jan 29 02:14:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	522157	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	799114	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	764579	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	360768	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	277059	50.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.54%	
35) Dibromofluoromethane	5.51	113	249868	48.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.66%	
50) Toluene-d8	8.71	98	986232	52.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.30%	
62) 4-Bromofluorobenzene	11.13	95	340525	52.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.30%	
Target Compounds						
16) Acetone	2.45	43	42045	15.585	ug/l	Qvalue 100
18) Methyl Acetate	2.78	43	9529	1.220	ug/l #	87
20) Methylene Chloride	2.86	84	38645	7.045	ug/l	97
25) 2-Butanone	4.70	43	10512	2.723	ug/l	97
43) Isopropyl Acetate	6.46	43	24385	2.075	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX012819\
Data File : VX007271.D
Acq On : 28 Jan 2019 19:52
Operator : JC/SP
Sample : K1254-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SOIL-1

Quant Time: Jan 29 02:14:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

