

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX012820\
 Data File : VX014746.D
 Acq On : 28 Jan 2020 13:04
 Operator : JC/SP
 Sample : L1245-03RE
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20200065-AMENDED-COMP-3RE

Quant Time: Jan 29 00:00:49 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 22 12:35:15 2020
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	430888	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	685114	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	628889	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	290202	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	255830	50.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.68%	
35) Dibromofluoromethane	5.48	113	124088	28.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	57.40%	
50) Toluene-d8	8.71	98	813678	48.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.92%	
62) 4-Bromofluorobenzene	11.14	95	281830	47.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.20%	
Target Compounds						
16) Acetone	2.43	43	54803	20.011	ug/l	97
20) Methylene Chloride	2.85	84	14061	2.755	ug/l	91


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX012820\
Data File : VX014746.D
Acq On : 28 Jan 2020 13:04
Operator : JC/SP
Sample : L1245-03RE
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20200065-AMENDED-COMP-3RE

Quant Time: Jan 29 00:00:49 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X012220W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 22 12:35:15 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.65 min Scan# 782

Delta R.T. -0.00 min

Lab File: VX014746.D

Acq: 28 Jan 2020 13:04

Instrument :

MSVOA_X

ClientSampleId :


20200065-AMENDED-COMP-3RE

Tot Ion:168 Resp: 430888

Ion Ratio Lower Upper




168 100

99 50.2 39.8 59.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#16

Acetone

Concen: 20.011 ug/l

RT: 2.43 min Scan# 253

Delta R.T. 0.00 min

Lab File: VX014746.D


Acq: 28 Jan 2020 13:04

Tgt Ion: 43 Resp: 54803

Ion Ratio Lower Upper



43 100


58 32.1 24.5 36.7



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0









#20
Methylene Chloride
Concen: 2.755 ug/l
RT: 2.85 min Scan# 322
Delta R.T. -0.00 min
Lab File: VX014746.D
Acq: 28 Jan 2020 13:04




Instrument :
MSVOA_X
ClientSampleId :
20200065-AMENDED-COMP-3RE


Tot Ion:	84	Resp:	14061
Ion	Ratio	Lower	Upper
84	100		
49	135.6	98.2	147.4
51	36.5	31.3	46.9
86	60.2	51.8	77.6



#33
1,2-Dichloroethane-d4
Concen: 50.843 ug/l
RT: 6.05 min Scan# 847
Delta R.T. -0.00 min
Lab File: VX014746.D
Acq: 28 Jan 2020 13:04

Tgt Ion:	65	Resp:	255830
Ion	Ratio	Lower	Upper
65	100		
67	52.5	0.0	107.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.85 min Scan# 978

Delta R.T. -0.00 min

Lab File: VX014746.D

Acq: 28 Jan 2020 13:04

Instrument :

MSVOA_X

ClientSampleId :

20200065-AMENDED-COMP-3RE


Tot Ion:114 Resp: 685114

Ion Ratio Lower Upper

114 100

63 20.9 0.0 40.6




88 15.4 0.0 31.2



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 28.704 ug/l

RT: 5.48 min Scan# 754

Delta R.T. -0.01 min

Lab File: VX014746.D

Acq: 28 Jan 2020 13:04


Tgt Ion:113 Resp: 124088

Ion Ratio Lower Upper

113 100

111 102.9 80.7 121.1



192 23.4 18.5 27.7




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 48.959 ug/l

RT: 8.71 min Scan# 1283

Delta R.T. -0.00 min

Lab File: VX014746.D

Acq: 28 Jan 2020 13:04

Instrument :

MSVOA_X

ClientSampleId :





20200065-AMENDED-COMP-3RE

Tot Ion: 98 Resp: 813678

Ion Ratio Lower Upper

98 100

100 65.7 53.0 79.6



#62

4-Bromofluorobenzene

Concen: 47.605 ug/l

RT: 11.14 min Scan# 1681

Delta R.T. -0.00 min

Lab File: VX014746.D

Acq: 28 Jan 2020 13:04




Tgt Ion: 95 Resp: 281830


Ion Ratio Lower Upper

95 100

174 88.8 0.0 176.0

176 85.7 0.0 172.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1513
Delta R.T. -0.00 min
Lab File: VX014746.D
Acq: 28 Jan 2020 13:04

Instrument :
MSVOA_X
ClientSampleId :
20200065-AMENDED-COMP-3RE

Tot Ion:117 Resp: 628889
Ion Ratio Lower Upper
117 100
82 52.9 43.2 64.8
119 31.9 25.8 38.6



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.07 min Scan# 1835
Delta R.T. -0.00 min
Lab File: VX014746.D
Acq: 28 Jan 2020 13:04

Tgt Ion:152 Resp: 290202
Ion Ratio Lower Upper
152 100
115 55.2 39.4 118.1
150 156.0 0.0 351.2

