

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012821\
 Data File : VX020787.D
 Acq On : 28 Jan 2021 09:37
 Operator : JC/SP
 Sample : VSTD0.135
 Misc : 25.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTD0.1305

Manual Integrations
 APPROVED

MMDadoda
 2/3/2021 9:42:26 AM

Quant Time: Jan 29 01:43:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXSIM012821.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jan 29 01:43:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)

Internal Standards					
1) 1,4-Difluorobenzene	6.823	114	12989	0.50 ug/L	0.00
5) Chlorobenzene-d5	10.089	117	11278	0.50 ug/L	0.00
11) 1,4-Dichlorobenzene-d4	12.063	152	5279	0.50 ug/L	0.00
System Monitoring Compounds					
2) Vinyl Chloride-d3	1.387	65	1021	0.09 ug/L	0.00
4) 1,2-Dichloroethane-d4	6.013	65	1172	0.10 ug/L	0.00
7) 1,2-Dichloropropane-d6	7.349	67	927	0.09 ug/L	-0.02
8) Toluene-d8	8.682	98	2661	0.09 ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	11.223	84	699m	0.08 ug/L	0.00
Target Compounds					
					Qvalue
3) Vinyl chloride	1.387	62	998	0.09 ug/L	89
6) Trichloroethene	7.180	95	834	0.08 ug/L	96
9) 1,2-Dibromoethane	9.656	107	490	0.08 ug/L #	90
12) 1,2,3-Trichloropropane	11.274	75	443	0.08 ug/L #	93
13) 1,2-Dibromo-3-chloropr...	12.971	75	92	0.09 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012821\
Data File : VX020787.D
Acq On : 28 Jan 2021 09:37
Operator : JC/SP
Sample : VSTD0.135
Misc : 25.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTD0.1305

Manual Integrations
APPROVED

MMDadoda
2/3/2021 9:42:26 AM

Quant Time: Jan 29 01:43:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXSIM012821.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jan 29 01:43:09 2021
Response via : Initial Calibration

