

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012821\
 Data File : VX020791.D
 Acq On : 28 Jan 2021 11:46
 Operator : JC/SP
 Sample : VSTDICV0.5
 Misc : 25.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VICV303

Quant Time: Jan 29 09:04:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXSIM012821.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jan 29 09:00:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.815	114	12143	0.50	ug/L	0.00
5) Chlorobenzene-d5	10.089	117	10763	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	12.062	152	4865	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.387	65	6061	0.56	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	112.00%
4) 1,2-Dichloroethane-d4	6.025	65	4816	0.45	ug/L	0.01
Spiked Amount	0.500	Range	70 - 130	Recovery	=	90.00%
7) 1,2-Dichloropropane-d6	7.361	67	5370	0.55	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	110.00%
8) Toluene-d8	8.681	98	13109	0.49	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	98.00%
10) 1,1,2,2-Tetrachloroeth...	11.222	84	4133	0.59	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	118.00%
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.387	62	5597	0.53	ug/L	100
6) Trichloroethene	7.173	95	4820	0.53	ug/L	97
9) 1,2-Dibromoethane	9.640	107	2888	0.54	ug/L	98
12) 1,2,3-Trichloropropane	11.274	75	2596	0.54	ug/L	97
13) 1,2-Dibromo-3-chloropr...	12.970	75	522	0.53	ug/L	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012821\
Data File : VX020791.D
Acq On : 28 Jan 2021 11:46
Operator : JC/SP
Sample : VSTDICV0.5
Misc : 25.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VICV303

Quant Time: Jan 29 09:04:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXSIM012821.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jan 29 09:00:02 2021
Response via : Initial Calibration

