

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012821\
 Data File : VX020788.D
 Acq On : 28 Jan 2021 10:04
 Operator : JC/SP
 Sample : VSTD0.536
 Misc : 25.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTD0.5306

Quant Time: Jan 29 01:44:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXSIM012821.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jan 29 01:43:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.822	114	12746	0.50	ug/L	0.00
5) Chlorobenzene-d5	10.093	117	11249	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	12.068	152	4898	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.387	65	6081	0.54	ug/L	0.00
4) 1,2-Dichloroethane-d4	6.013	65	4952	0.42	ug/L	0.00
7) 1,2-Dichloropropane-d6	7.368	67	5465	0.51	ug/L	0.00
8) Toluene-d8	8.685	98	13593	0.47	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	11.228	84	3812	0.46	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.387	62	5698	0.54	ug/L	100
6) Trichloroethene	7.180	95	5004	0.47	ug/L	100
9) 1,2-Dibromoethane	9.644	107	2972	0.49	ug/L	100
12) 1,2,3-Trichloropropane	11.262	75	2745	0.52	ug/L	100
13) 1,2-Dibromo-3-chloropr...	12.976	75	520	0.52	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012821\
Data File : VX020788.D
Acq On : 28 Jan 2021 10:04
Operator : JC/SP
Sample : VSTD0.536
Misc : 25.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTD0.5306

Quant Time: Jan 29 01:44:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXSIM012821.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jan 29 01:43:09 2021
Response via : Initial Calibration

