

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012821\
 Data File : VX020792.D
 Acq On : 28 Jan 2021 12:18
 Operator : JC/SP
 Sample : VX0128WBL01
 Misc : 25.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK302

Quant Time: Jan 29 09:05:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXSIM012821.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jan 29 09:00:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.822	114	11878	0.50	ug/L	0.00
5) Chlorobenzene-d5	10.092	117	10361	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	12.068	152	4492	0.50	ug/L	0.00

System Monitoring Compounds						
2) Vinyl Chloride-d3	1.387	65	5783	0.54	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	108.00%
4) 1,2-Dichloroethane-d4	6.013	65	4842	0.47	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	94.00%
7) 1,2-Dichloropropane-d6	7.349	67	5221	0.56	ug/L	-0.02
Spiked Amount	0.500	Range	60 - 140	Recovery	=	112.00%
8) Toluene-d8	8.685	98	12683	0.50	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	100.00%
10) 1,1,2,2-Tetrachloroeth...	11.228	84	3732	0.55	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	110.00%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012821\
Data File : VX020792.D
Acq On : 28 Jan 2021 12:18
Operator : JC/SP
Sample : VX0128WBL01
Misc : 25.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK302

Quant Time: Jan 29 09:05:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXSIM012821.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jan 29 09:00:02 2021
Response via : Initial Calibration

