

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012821\
 Data File : VX020795.D
 Acq On : 28 Jan 2021 15:21
 Operator : JC/SP
 Sample : VSTDCCC0.5EC
 Misc : 25.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTD0.5303

Quant Time: Jan 29 09:08:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXSIM012821.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jan 29 09:00:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.823	114	11133	0.50	ug/L	0.00
5) Chlorobenzene-d5	10.089	117	10061	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	12.063	152	4652	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.387	65	5842	0.58	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	116.00%
4) 1,2-Dichloroethane-d4	6.013	65	4697	0.48	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	96.00%
7) 1,2-Dichloropropane-d6	7.350	67	5228	0.58	ug/L	-0.02
Spiked Amount	0.500	Range	60 - 140	Recovery	=	116.00%
8) Toluene-d8	8.682	98	12242	0.49	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	98.00%
10) 1,1,2,2-Tetrachloroeth...	11.223	84	4087	0.62	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	124.00%#
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.387	62	5264	0.55	ug/L	100
6) Trichloroethene	7.180	95	4536	0.54	ug/L	97
9) 1,2-Dibromoethane	9.641	107	2670	0.53	ug/L	100
12) 1,2,3-Trichloropropane	11.274	75	2429	0.53	ug/L	98
13) 1,2-Dibromo-3-chloropr...	12.971	75	494	0.52	ug/L	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012821\
Data File : VX020795.D
Acq On : 28 Jan 2021 15:21
Operator : JC/SP
Sample : VSTDCCC0.5EC
Misc : 25.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTD0.5303

Quant Time: Jan 29 09:08:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXSIM012821.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jan 29 09:00:02 2021
Response via : Initial Calibration

