

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012822\
 Data File : VX026646.D
 Acq On : 28 Jan 2022 13:16
 Operator : JC/MD
 Sample : N1297-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-17

Quant Time: Jan 29 02:18:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	124168	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	208964	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	177571	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	71288	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	80014	46.130	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.260%
35) Dibromofluoromethane	5.391	113	65959	48.879	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.760%
50) Toluene-d8	8.653	98	241676	49.699	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.400%
62) 4-Bromofluorobenzene	11.079	95	81058	46.924	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.840%
Target Compounds						
30) Chloroform	5.099	83	2082	0.766	ug/l	Qvalue 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012822\
Data File : VX026646.D
Acq On : 28 Jan 2022 13:16
Operator : JC/MD
Sample : N1297-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-17

Quant Time: Jan 29 02:18:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

