

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012822\
 Data File : VX026649.D
 Acq On : 28 Jan 2022 14:26
 Operator : JC/MD
 Sample : N1297-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB012522

Quant Time: Jan 29 02:19:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	126553	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	215571	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	186654	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	75518	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	82871	46.877	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.760%
35) Dibromofluoromethane	5.391	113	67733	48.656	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.320%
50) Toluene-d8	8.653	98	251026	50.040	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.080%
62) 4-Bromofluorobenzene	11.085	95	86147	48.342	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012822\
Data File : VX026649.D
Acq On : 28 Jan 2022 14:26
Operator : JC/MD
Sample : N1297-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB012522

Quant Time: Jan 29 02:19:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

