

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012822\
 Data File : VX026650.D
 Acq On : 28 Jan 2022 14:49
 Operator : JC/MD
 Sample : N1297-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB012622

Quant Time: Jan 29 02:19:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	124903	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	213552	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	178888	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	68981	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	81903	46.941	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.880%
35) Dibromofluoromethane	5.391	113	66079	47.916	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.840%
50) Toluene-d8	8.653	98	243095	48.917	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.840%
62) 4-Bromofluorobenzene	11.079	95	80286	45.479	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.960%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	1451	1.631	ug/l	94
30) Chloroform	5.093	83	902	0.330	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012822\
Data File : VX026650.D
Acq On : 28 Jan 2022 14:49
Operator : JC/MD
Sample : N1297-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB012622

Quant Time: Jan 29 02:19:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

