

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012822\
 Data File : VX026652.D
 Acq On : 28 Jan 2022 15:35
 Operator : JC/MD
 Sample : N1323-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Jan 29 02:19:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	119164	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	204890	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	173103	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	70838	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	79739	47.902	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.800%
35) Dibromofluoromethane	5.391	113	65032	49.151	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.300%
50) Toluene-d8	8.653	98	234449	49.172	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.340%
62) 4-Bromofluorobenzene	11.079	95	79291	46.814	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.620%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012822\
Data File : VX026652.D
Acq On : 28 Jan 2022 15:35
Operator : JC/MD
Sample : N1323-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Jan 29 02:19:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

