

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012822\
 Data File : VX026654.D
 Acq On : 28 Jan 2022 16:21
 Operator : JC/MD
 Sample : N1323-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Jan 29 02:19:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	119040	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	203370	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	175347	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	73371	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	79070	47.549	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.100%
35) Dibromofluoromethane	5.391	113	65008	49.500	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.000%
50) Toluene-d8	8.653	98	234041	49.453	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.900%
62) 4-Bromofluorobenzene	11.085	95	80627	47.958	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012822\
Data File : VX026654.D
Acq On : 28 Jan 2022 16:21
Operator : JC/MD
Sample : N1323-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Jan 29 02:19:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

