

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012822\
 Data File : VX026657.D
 Acq On : 28 Jan 2022 17:31
 Operator : JC/MD
 Sample : N1297-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2EA

Quant Time: Jan 29 02:20:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	116636	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	199519	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	172158	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	71822	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	77854	47.783	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.560%
35) Dibromofluoromethane	5.391	113	63056	48.940	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.880%
50) Toluene-d8	8.653	98	228210	49.152	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.300%
62) 4-Bromofluorobenzene	11.085	95	80149	48.594	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.180%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	2060	2.479	ug/l	96
30) Chloroform	5.105	83	7516	2.944	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012822\
Data File : VX026657.D
Acq On : 28 Jan 2022 17:31
Operator : JC/MD
Sample : N1297-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2EA

Quant Time: Jan 29 02:20:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

