

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012822\
 Data File : VX026658.D
 Acq On : 28 Jan 2022 17:54
 Operator : JC/MD
 Sample : N1297-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2EB

Quant Time: Jan 29 02:20:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	120174	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	206460	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	177037	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	75373	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	79389	47.291	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.580%
35) Dibromofluoromethane	5.391	113	65270	48.955	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.920%
50) Toluene-d8	8.653	98	236010	49.123	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.240%
62) 4-Bromofluorobenzene	11.079	95	82512	48.345	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.700%
Target Compounds						
30) Chloroform	5.099	83	7649	2.907	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012822\
Data File : VX026658.D
Acq On : 28 Jan 2022 17:54
Operator : JC/MD
Sample : N1297-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2EB

Quant Time: Jan 29 02:20:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

