

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012822\
 Data File : VX026659.D
 Acq On : 28 Jan 2022 18:17
 Operator : JC/MD
 Sample : N1297-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-11

Quant Time: Jan 29 02:20:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	121352	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	209362	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	182074	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76090	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	81138	47.863	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.720%
35) Dibromofluoromethane	5.391	113	66521	49.202	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.400%
50) Toluene-d8	8.653	98	242869	49.850	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.700%
62) 4-Bromofluorobenzene	11.085	95	83565	48.283	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.560%
Target Compounds						
64) Tetrachloroethene	9.274	164	2080	1.209	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012822\
Data File : VX026659.D
Acq On : 28 Jan 2022 18:17
Operator : JC/MD
Sample : N1297-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-11

Quant Time: Jan 29 02:20:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

