

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012822\
 Data File : VX026661.D
 Acq On : 28 Jan 2022 19:03
 Operator : JC/MD
 Sample : N1297-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GES-2

Quant Time: Jan 29 02:20:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	123129	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	207709	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	182597	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	75536	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	80010	46.517	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.040%
35) Dibromofluoromethane	5.391	113	64698	48.235	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.460%
50) Toluene-d8	8.653	98	241470	49.957	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.920%
62) 4-Bromofluorobenzene	11.079	95	83704	48.749	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.500%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	1758	2.004	ug/l	94
39) Methylcyclohexane	7.385	83	5612	2.197	ug/l	91
52) Toluene	8.726	92	863	0.240	ug/l	83
64) Tetrachloroethene	9.275	164	1804	1.046	ug/l	95
67) Ethyl Benzene	10.195	91	5029	0.720	ug/l	98
68) m/p-Xylenes	10.305	106	3794	1.451	ug/l	96
69) o-Xylene	10.646	106	3471	1.376	ug/l	98
78) n-propylbenzene	11.305	91	1446	0.205	ug/l	98
80) 1,3,5-Trimethylbenzene	11.451	105	3606	0.706	ug/l	98
84) 1,2,4-Trimethylbenzene	11.756	105	7480	1.465	ug/l	96
89) n-Butylbenzene	12.329	91	2729	0.615	ug/l #	42
95) Naphthalene	13.780	128	6266	1.166	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012822\
Data File : VX026661.D
Acq On : 28 Jan 2022 19:03
Operator : JC/MD
Sample : N1297-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GES-2

Quant Time: Jan 29 02:20:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

