

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012822\
 Data File : VX026665.D
 Acq On : 28 Jan 2022 20:36
 Operator : JC/MD
 Sample : N1319-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-1D

Quant Time: Jan 29 02:21:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	128806	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	220962	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	188457	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	80054	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	83977	46.671	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.340%
35) Dibromofluoromethane	5.391	113	69048	48.390	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.780%
50) Toluene-d8	8.653	98	253470	49.295	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.580%
62) 4-Bromofluorobenzene	11.079	95	87635	47.977	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.960%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	13061	14.234	ug/l	99
25) 2-Butanone	4.574	43	2596	2.175	ug/l	95
43) Isopropyl Acetate	6.348	43	11939	3.244	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012822\
Data File : VX026665.D
Acq On : 28 Jan 2022 20:36
Operator : JC/MD
Sample : N1319-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-1D

Quant Time: Jan 29 02:21:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

