

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012822\
 Data File : VX026666.D
 Acq On : 28 Jan 2022 20:59
 Operator : JC/MD
 Sample : N1319-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-2S

Quant Time: Jan 29 02:21:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	129272	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	220581	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	190989	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	81549	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	84474	46.778	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.560%
35) Dibromofluoromethane	5.391	113	69669	48.910	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.820%
50) Toluene-d8	8.653	98	253908	49.465	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.920%
62) 4-Bromofluorobenzene	11.079	95	89386	49.020	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.040%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	12526	13.602	ug/l	100
25) 2-Butanone	4.574	43	2658	2.219	ug/l	96
43) Isopropyl Acetate	6.342	43	11916	3.243	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012822\
Data File : VX026666.D
Acq On : 28 Jan 2022 20:59
Operator : JC/MD
Sample : N1319-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-2S

Quant Time: Jan 29 02:21:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

