

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX012822\
 Data File : VX026667.D
 Acq On : 28 Jan 2022 21:22
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 29 02:22:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	87	0.00
2 T	Dichlorodifluoromethane	50.000	41.248	17.5	71	0.00
3 P	Chloromethane	50.000	41.434	17.1	74	0.00
4 C	Vinyl Chloride	50.000	44.800	10.4#	76	0.00
5 T	Bromomethane	50.000	38.034	23.9	67	0.00
6 T	Chloroethane	50.000	43.852	12.3	77	0.00
7 T	Trichlorofluoromethane	50.000	44.780	10.4	75	0.00
8 T	Diethyl Ether	50.000	48.455	3.1	81	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	49.110	1.8	83	0.00
10 T	Methyl Iodide	50.000	48.217	3.6	86	0.00
11 T	Tert butyl alcohol	250.000	259.739	-3.9	86	-0.01
12 CM	1,1-Dichloroethene	50.000	49.153	1.7#	83	0.00
13 T	Acrolein	250.000	257.152	-2.9	94	0.00
14 T	Allyl chloride	50.000	50.125	-0.3	83	0.00
15 T	Acrylonitrile	250.000	262.497	-5.0	90	0.00
16 T	Acetone	250.000	181.977	27.2#	61	0.00
17 T	Carbon Disulfide	50.000	41.559	16.9	73	0.00
18 T	Methyl Acetate	50.000	51.806	-3.6	90	0.00
19 T	Methyl tert-butyl Ether	50.000	50.186	-0.4	85	0.00
20 T	Methylene Chloride	50.000	48.978	2.0	87	0.00
21 T	trans-1,2-Dichloroethene	50.000	49.286	1.4	83	0.00
22 T	Diisopropyl ether	50.000	50.949	-1.9	86	0.00
23 T	Vinyl Acetate	250.000	261.871	-4.7	86	0.00
24 P	1,1-Dichloroethane	50.000	50.526	-1.1	85	0.00
25 T	2-Butanone	250.000	231.613	7.4	77	0.00
26 T	2,2-Dichloropropane	50.000	47.948	4.1	78	0.00
27 T	cis-1,2-Dichloroethene	50.000	51.902	-3.8	86	0.00
28 T	Bromochloromethane	50.000	49.125	1.8	82	0.00
29 T	Tetrahydrofuran	250.000	260.607	-4.2	89	0.00
30 C	Chloroform	50.000	50.109	-0.2#	85	0.00
31 T	Cyclohexane	50.000	46.257	7.5	80	0.00
32 T	1,1,1-Trichloroethane	50.000	49.961	0.1	82	0.00
33 S	1,2-Dichloroethane-d4	50.000	46.950	6.1	79	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	88	0.00
35 S	Dibromofluoromethane	50.000	51.287	-2.6	86	0.00
36 T	1,1-Dichloropropene	50.000	47.343	5.3	81	0.00
37 T	Ethyl Acetate	50.000	51.142	-2.3	86	0.00
38 T	Carbon Tetrachloride	50.000	49.637	0.7	82	0.00
39 T	Methylcyclohexane	50.000	47.587	4.8	80	0.00
40 TM	Benzene	50.000	50.567	-1.1	86	0.00
41 T	Methacrylonitrile	50.000	53.280	-6.6	89	0.00
42 TM	1,2-Dichloroethane	50.000	47.929	4.1	81	0.00
43 T	Isopropyl Acetate	50.000	52.633	-5.3	87	0.00
44 TM	Trichloroethene	50.000	50.985	-2.0	86	0.00
45 C	1,2-Dichloropropane	50.000	52.465	-4.9#	88	0.00
46 T	Dibromomethane	50.000	51.198	-2.4	86	0.00
47 T	Bromodichloromethane	50.000	51.182	-2.4	84	0.00
48 T	Methyl methacrylate	50.000	51.909	-3.8	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX012822\
 Data File : VX026667.D
 Acq On : 28 Jan 2022 21:22
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 29 02:22:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1029.747	-3.0	85	0.00
50 S	Toluene-d8	50.000	50.032	-0.1	85	0.00
51 T	4-Methyl-2-Pentanone	250.000	262.440	-5.0	88	0.00
52 CM	Toluene	50.000	50.848	-1.7#	86	0.00
53 T	t-1,3-Dichloropropene	50.000	48.238	3.5	86	0.00
54 T	cis-1,3-Dichloropropene	50.000	53.297	-6.6	85	0.00
55 T	1,1,2-Trichloroethane	50.000	52.788	-5.6	89	0.00
56 T	Ethyl methacrylate	50.000	54.629	-9.3	88	0.00
57 T	1,3-Dichloropropane	50.000	51.945	-3.9	88	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	256.945	-2.8	85	0.00
59 T	2-Hexanone	250.000	251.700	-0.7	83	0.00
60 T	Dibromochloromethane	50.000	53.701	-7.4	87	0.00
61 T	1,2-Dibromoethane	50.000	52.160	-4.3	87	0.00
62 S	4-Bromofluorobenzene	50.000	50.068	-0.1	84	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	88	0.00
64 T	Tetrachloroethene	50.000	51.763	-3.5	88	0.00
65 PM	Chlorobenzene	50.000	51.291	-2.6	88	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	52.783	-5.6	88	0.00
67 C	Ethyl Benzene	50.000	51.117	-2.2#	85	0.00
68 T	m/p-Xylenes	100.000	103.565	-3.6	87	0.00
69 T	o-Xylene	50.000	52.273	-4.5	87	0.00
70 T	Styrene	50.000	53.518	-7.0	87	0.00
71 P	Bromoform	50.000	49.122	1.8	89	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	88	0.00
73 T	Isopropylbenzene	50.000	50.286	-0.6	86	0.00
74 T	N-amyl acetate	50.000	55.575	-11.2	89	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	53.614	-7.2	92	0.00
76 T	1,2,3-Trichloropropane	50.000	52.937	-5.9	89	0.00
77 T	Bromobenzene	50.000	52.195	-4.4	89	0.00
78 T	n-propylbenzene	50.000	50.604	-1.2	86	0.00
79 T	2-Chlorotoluene	50.000	49.446	1.1	86	0.00
80 T	1,3,5-Trimethylbenzene	50.000	50.792	-1.6	86	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	50.248	-0.5	92	0.00
82 T	4-Chlorotoluene	50.000	49.083	1.8	84	0.00
83 T	tert-Butylbenzene	50.000	51.716	-3.4	87	0.00
84 T	1,2,4-Trimethylbenzene	50.000	50.577	-1.2	85	0.00
85 T	sec-Butylbenzene	50.000	51.873	-3.7	87	0.00
86 T	p-Isopropyltoluene	50.000	51.573	-3.1	86	0.00
87 T	1,3-Dichlorobenzene	50.000	51.167	-2.3	90	0.00
88 T	1,4-Dichlorobenzene	50.000	50.693	-1.4	88	0.00
89 T	n-Butylbenzene	50.000	50.264	-0.5	83	0.00
90 T	Hexachloroethane	50.000	48.306	3.4	87	0.00
91 T	1,2-Dichlorobenzene	50.000	52.956	-5.9	90	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	53.457	-6.9	87	0.00
93 T	1,2,4-Trichlorobenzene	50.000	51.396	-2.8	86	0.00
94 T	Hexachlorobutadiene	50.000	49.872	0.3	86	0.00
95 T	Naphthalene	50.000	55.481	-11.0	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012822\
Data File : VX026667.D
Acq On : 28 Jan 2022 21:22
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jan 29 02:22:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	53.876	-7.8	91	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6