

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012825\
 Data File : VX044747.D
 Acq On : 28 Jan 2025 16:21
 Operator : JC/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jan 29 07:41:03 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML012825WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jan 28 12:04:44 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	305709	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.049	117	257994	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.018	152	103662	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	107696	46.800	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.600%
7) Chloroethane-d5	1.642	69	29746	34.026	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	68.060%#
11) 1,1-Dichloroethene-d2	2.294	65	46925	47.706	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.420%
21) 2-Butanone-d5	4.458	46	120996	122.992	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	122.990%
24) Chloroform-d	5.050	84	186973	46.238	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.480%
26) 1,2-Dichloroethane-d4	5.952	65	117326	48.338	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.680%
32) Benzene-d6	5.964	84	401328	49.750	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.500%
36) 1,2-Dichloropropane-d6	7.306	67	126815	49.666	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.340%
41) Toluene-d8	8.647	98	350341	49.069	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.140%
43) trans-1,3-Dichloroprop...	8.945	79	57967	47.861	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.720%
47) 2-Hexanone-d5	9.384	63	74816	99.005	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	99.010%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	158392	49.920	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.840%
66) 1,2-Dichlorobenzene-d4	12.317	152	102416	50.493	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.980%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012825\
Data File : VX044747.D
Acq On : 28 Jan 2025 16:21
Operator : JC/MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Jan 29 07:41:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML012825WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jan 28 12:04:44 2025
Response via : Initial Calibration

