

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX012919\
 Data File : VX007300.D
 Acq On : 29 Jan 2019 20:35
 Operator : JC/SP
 Sample : K1265-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 26 Sample Multiplier: 1


Quant Time: Jan 30 00:48:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	519137	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	798273	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	766467	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	383775	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	278441	51.32	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.64%	
35) Dibromofluoromethane	5.51	113	250648	49.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.06%	
50) Toluene-d8	8.71	98	974281	52.07	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.14%	
62) 4-Bromofluorobenzene	11.13	95	342104	52.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.88%	
Target Compounds						
					Qvalue	
8) Diethyl Ether	2.20	74	1894	0.552	ug/l	74
11) Tert butyl alcohol	3.07	59	1737	1.357	ug/l #	53
13) Acrolein	2.31	56	1478	5.729	ug/l #	64
16) Acetone	2.45	43	216101	80.570	ug/l	99
17) Carbon Disulfide	2.57	76	5075	0.413	ug/l	99
18) Methyl Acetate	2.78	43	19515	2.512	ug/l	94
20) Methylene Chloride	2.86	84	755342	149.186	ug/l	96
25) 2-Butanone	4.70	43	60433	15.747	ug/l	97
29) Tetrahydrofuran	5.16	42	2856	1.145	ug/l	95
36) 1,1-Dichloropropene	5.67	75	34309	4.943	ug/l #	49
37) Ethyl Acetate	4.70	43	60433	8.122	ug/l #	67
38) Carbon Tetrachloride	5.66	117	47323	6.777	ug/l #	15
43) Isopropyl Acetate	6.46	43	30263	2.579	ug/l	98
48) Methyl methacrylate	7.69	41	18482	3.066	ug/l #	12
49) 1,4-Dioxane	7.76	88	367	Below Cal	#	43
51) 4-Methyl-2-Pentanone	8.62	43	142141	18.845	ug/l #	46
52) Toluene	8.79	92	219547	16.130	ug/l	97
54) cis-1,3-Dichloropropene	8.62	75	65481	8.396	ug/l #	64
57) 1,3-Dichloropropane	9.54	76	14764	1.654	ug/l #	45
59) 2-Hexanone	9.50	43	6269	1.091	ug/l	99
68) m/p-Xylenes	10.36	106	3403	0.313	ug/l	96
69) o-Xylene	10.69	106	2863	0.270	ug/l	97
76) 1,2,3-Trichloropropane	11.13	75	164977	21.213	ug/l #	40
80) 1,3,5-Trimethylbenzene	11.51	105	8833	0.391	ug/l	94
81) trans-1,4-Dichloro-2-buten	11.13	75	164977	62.958	ug/l #	10
84) 1,2,4-Trimethylbenzene	11.80	105	22905	1.007	ug/l	94
86) p-Isopropyltoluene	12.06	119	13179	0.547	ug/l	89
90) Hexachloroethane	12.58	117	936	3.137	ug/l #	11
95) Naphthalene	13.83	128	19185	0.694	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX012919\
Data File : VX007300.D
Acq On : 29 Jan 2019 20:35
Operator : JC/SP
Sample : K1265-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 26 Sample Multiplier: 1

Quant Time: Jan 30 00:16:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 740

Delta R.T. -0.00 min

Lab File: VX007300.D


Acq: 29 Jan 2019 20:35

Tot Ion:168 Resp: 519137

Ion Ratio Lower Upper

168 100

99 48.5 41.8 62.8



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.67

200000



150000

100000

50000

0

Time-->



#8

Diethyl Ether

Concen: 0.552 ug/l

RT: 2.20 min Scan# 171

Delta R.T. 0.01 min

Lab File: VX007300.D


Acq: 29 Jan 2019 20:35

Tgt Ion: 74 Resp: 1894

Ion Ratio Lower Upper

74 100

45 117.1 46.3 138.9



Abundance Ion 74.00 (73.70 to 74.70): VX0

Ion 45.00 (44.70 to 45.70): VX0

2.20

2000


1500


1000

500

0

Time-->









#11

Tert butyl alcohol
 Concen: 1.357 ug/l
 RT: 3.07 min Scan# 314
 Delta R.T. -0.01 min
 Lab File: VX007300.D
 Acq: 29 Jan 2019 20:35




Tot Ion: 59 Resp: 1737
 Ion Ratio Lower Upper
 59 100
 57 28.1 8.5 12.7#




#13

Acrolein
 Concen: 5.729 ug/l
 RT: 2.31 min Scan# 189
 Delta R.T. 0.01 min
 Lab File: VX007300.D
 Acq: 29 Jan 2019 20:35

Tgt Ion: 56 Resp: 1478
 Ion Ratio Lower Upper
 56 100
 55 42.7 58.2 87.4#





#16

Acetone

Concen: 80.570 ug/l

RT: 2.45 min Scan# 212


Delta R.T. -0.00 min

Lab File: VX007300.D

Acq: 29 Jan 2019 20:35

Tot Ion: 43 Resp: 216101

Ion	Ratio	Lower	Upper
43	100		
58	31.9	25.0	37.6



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0

2.45

150000



100000

50000

0

Time-->

2.30 2.40 2.50 2.60



#17

Carbon Disulfide

Concen: 0.413 ug/l

RT: 2.57 min Scan# 232


Delta R.T. -0.00 min

Lab File: VX007300.D

Acq: 29 Jan 2019 20:35

Tgt Ion: 76 Resp: 5075

Ion	Ratio	Lower	Upper
76	100		
78	9.1	7.0	10.6



Abundance Ion 75.90 (75.60 to 76.60): VX0

Ion 77.90 (77.60 to 78.60): VX0

2.57

3000


2000


1000

0

Time-->





2.50 2.55 2.60 2.65








#18
Methvl Acetate
Concen: 2.512 ug/l
RT: 2.78 min Scan# 267
Delta R.T. -0.00 min
Lab File: VX007300.D
Acq: 29 Jan 2019 20:35


Tot Ion: 43 Resp: 19515
Ion Ratio Lower Upper
43 100
74 26.4 18.6 28.0



#20
Methylene Chloride
Concen: 149.186 ug/l
RT: 2.86 min Scan# 280
Delta R.T. 0.00 min
Lab File: VX007300.D
Acq: 29 Jan 2019 20:35

Tgt Ion: 84 Resp: 755342
Ion Ratio Lower Upper
84 100
49 120.3 91.8 137.6
51 36.4 28.5 42.7
86 65.3 50.2 75.2





#25

2-Butanone

Concen: 15.747 ug/l

RT: 4.70 min Scan# 582

Delta R.T. -0.00 min

Lab File: VX007300.D





Acq: 29 Jan 2019 20:35

Tot Ion: 43 Resp: 60433

Ion Ratio Lower Upper

43 100

72 29.7 22.4 33.6



#29

Tetrahydrofuran

Concen: 1.145 ug/l

RT: 5.16 min Scan# 657

Delta R.T. 0.01 min

Lab File: VX007300.D

Acq: 29 Jan 2019 20:35




Tgt Ion: 42 Resp: 2856


Ion Ratio Lower Upper

42 100

72 55.1 38.9 58.3

71 45.9 36.2 54.4





#33

1,2-Dichloroethane-d4

Concen: 51.322 ug/l

RT: 6.07 min Scan# 806

Delta R.T. -0.00 min

Lab File: VX007300.D





Acq: 29 Jan 2019 20:35

Tot Ion: 65 Resp: 278441

Ion Ratio Lower Upper

65 100

67 53.6 0.0 105.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.86 min Scan# 936

Delta R.T. -0.00 min

Lab File: VX007300.D

Acq: 29 Jan 2019 20:35




Tgt Ion: 114 Resp: 798273


Ion Ratio Lower Upper

114 100

63 18.3 0.0 34.8

88 15.1 0.0 28.8





#35

Dibromofluoromethane

Concen: 49.030 ug/l

RT: 5.51 min Scan# 714

Delta R.T. 0.01 min

Lab File: VX007300.D

Acq: 29 Jan 2019 20:35


Tot Ion:113 Resp: 250648

Ion Ratio Lower Upper




113 100

111 102.4 81.6 122.4

192 22.2 16.7 25.1



Abundance Ion 112.80 (112.50 to 113.50): V
Ion 110.80 (110.50 to 111.50): V
Ion 191.70 (191.40 to 192.40): V



#36

1,1-Dichloropropene

Concen: 4.943 ug/l

RT: 5.67 min Scan# 740

Delta R.T. -0.13 min

Lab File: VX007300.D

Acq: 29 Jan 2019 20:35


Tgt Ion: 75 Resp: 34309

Ion Ratio Lower Upper



75 100


110 11.1 20.2 60.5#

77 0.7 24.8 37.2#



Abundance Ion 74.90 (74.60 to 75.60): VX0
Ion 109.90 (109.60 to 110.60): V
Ion 76.90 (76.60 to 77.60): VX0





#37

Ethyl Acetate

Concen: 8.122 ug/l

RT: 4.70 min Scan# 582


Delta R.T. -0.15 min

Lab File: VX007300.D

Acq: 29 Jan 2019 20:35

Tot Ion: 43 Resp: 60433




Ion	Ratio	Lower	Upper
43	100		
61	0.0	11.9	17.9#
70	0.9	9.7	14.5#



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 60.90 (60.60 to 61.60): VX0

Ion 70.00 (69.70 to 70.70): VX0



#38

Carbon Tetrachloride

Concen: 6.777 ug/l

RT: 5.66 min Scan# 739


Delta R.T. -0.12 min

Lab File: VX007300.D

Acq: 29 Jan 2019 20:35

Tgt Ion: 117 Resp: 47323



Ion	Ratio	Lower	Upper
117	100		
119	5.4	79.4	119.2#
121	0.0	24.0	36.0#




Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V





#43

Isopropyl Acetate

Concen: 2.579 ug/l

RT: 6.46 min Scan# 871

Delta R.T. 0.01 min

Lab File: VX007300.D

Acq: 29 Jan 2019 20:35


Tot Ion: 43 Resp: 30263

Ion Ratio Lower Upper

43 100

61 19.2 16.8 25.2




87 15.7 12.4 18.6



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 61.00 (60.70 to 61.70): VX0

Ion 87.00 (86.70 to 87.70): VX0



#48

Methyl methacrylate

Concen: 3.066 ug/l

RT: 7.69 min Scan# 1072

Delta R.T. -0.09 min

Lab File: VX007300.D

Acq: 29 Jan 2019 20:35


Tgt Ion: 41 Resp: 18482

Ion Ratio Lower Upper

41 100

69 0.4 72.4 108.6#



39 0.0 45.8 68.8#




Abundance Ion 41.00 (40.70 to 41.70): VX0

Ion 69.00 (68.70 to 69.70): VX0

Ion 39.00 (38.70 to 39.70): VX0





#49

1,4-Dioxane

Concen: Below Cal

RT: 7.76 min Scan# 1083


Delta R.T. 0.01 min

Lab File: VX007300.D

Acq: 29 Jan 2019 20:35

Tot Ion: 88 Resp: 367

Ion	Ratio	Lower	Upper
88	100		
43	0.0	22.2	33.4#
58	109.5	51.0	76.4#






Abundance

Ion 88.00 (87.70 to 88.70): VX0

Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0



#50

Toluene-d8

Concen: 52.068 ug/l

RT: 8.71 min Scan# 1240


Delta R.T. -0.00 min

Lab File: VX007300.D

Acq: 29 Jan 2019 20:35

Tgt Ion: 98 Resp: 974281



Ion	Ratio	Lower	Upper
98	100		
100	64.2	52.0	78.0




Abundance

Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): VX0





#51

4-Methyl-2-Pentanone

Concen: 18.845 ug/l

RT: 8.62 min Scan# 1224

Delta R.T. -0.03 min

Lab File: VX007300.D





Acq: 29 Jan 2019 20:35

Tot Ion: 43 Resp: 142141

Ion Ratio Lower Upper

43 100

58 6.7 31.8 47.6#



#52

Toluene

Concen: 16.130 ug/l

RT: 8.79 min Scan# 1252

Delta R.T. -0.00 min

Lab File: VX007300.D




Acq: 29 Jan 2019 20:35


Tgt Ion: 92 Resp: 219547

Ion Ratio Lower Upper

92 100

91 175.5 137.6 206.4





#54

cis-1,3-Dichloropropene

Concen: 8.396 ug/l

RT: 8.62 min Scan# 1224


Delta R.T. 0.18 min

Lab File: VX007300.D

Acq: 29 Jan 2019 20:35

Tot Ion: 75 Resp: 65481




Ion	Ratio	Lower	Upper
75	100		
77	0.5	26.2	39.2#
39	31.8	36.7	55.1#



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 76.90 (76.60 to 77.60): VX0

Ion 39.00 (38.70 to 39.70): VX0



#57

1,3-Dichloropropene

Concen: 1.654 ug/l

RT: 9.54 min Scan# 1376


Delta R.T. 0.17 min

Lab File: VX007300.D

Acq: 29 Jan 2019 20:35



Tgt Ion: 76 Resp: 14764


Ion	Ratio	Lower	Upper
76	100		
78	1.8	26.0	39.0#



Abundance Ion 75.90 (75.60 to 76.60): VX0

Ion 77.90 (77.60 to 78.60): VX0





#59

2-Hexanone

Concen: 1.091 ug/l

RT: 9.50 min Scan# 1369

Delta R.T. 0.01 min

Lab File: VX007300.D





Acq: 29 Jan 2019 20:35

Tot Ion: 43 Resp: 6269

Ion Ratio Lower Upper

43 100

58 55.0 27.7 83.1



#62

4-Bromofluorobenzene

Concen: 52.443 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. -0.00 min

Lab File: VX007300.D

Acq: 29 Jan 2019 20:35




Tgt Ion: 95 Resp: 342104


Ion Ratio Lower Upper

95 100

174 94.3 0.0 182.2


176 90.5 0.0 178.8








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.12 min Scan# 1470
Delta R.T. -0.00 min
Lab File: VX007300.D
Acq: 29 Jan 2019 20:35

Tot Ion	Ratio	Lower	Upper
117	100		
82	49.5	41.0	61.6
119	31.7	25.4	38.0




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V






#68
m/p-Xylenes
Concen: 0.313 ug/l
RT: 10.36 min Scan# 1510
Delta R.T. -0.00 min
Lab File: VX007300.D
Acq: 29 Jan 2019 20:35

Tgt Ion	Ratio	Lower	Upper
106	100		
91	201.2	155.8	233.6



Abundance Ion 106.00 (105.70 to 106.70): V
Ion 91.00 (90.70 to 91.70): VX0





#69

o-Xylene

Concen: 0.270 ug/l

RT: 10.69 min Scan# 1565

Delta R.T. -0.01 min

Lab File: VX007300.D


Acq: 29 Jan 2019 20:35

Tot Ion:106 Resp: 2863

Ion Ratio Lower Upper




106 100

91 198.4 101.3 303.7



Abundance Ion 106.00 (105.70 to 106.70): V

Ion 91.00 (90.70 to 91.70): VX0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.08 min Scan# 1792

Delta R.T. -0.00 min

Lab File: VX007300.D

Acq: 29 Jan 2019 20:35


Tgt Ion:152 Resp: 383775

Ion Ratio Lower Upper

152 100

115 54.2 39.1 117.2



150 155.7 0.0 344.8




Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V









#76

1,2,3-Trichloropropane
Concen: 21.213 ug/l
RT: 11.13 min Scan# 1637
Delta R.T. -0.16 min
Lab File: VX007300.D
Acq: 29 Jan 2019 20:35




Tot Ion: 75 Resp: 164977
Ion Ratio Lower Upper
75 100
77 1.5 18.5 55.5#




#80

1,3,5-Trimethylbenzene
Concen: 0.391 ug/l
RT: 11.51 min Scan# 1698
Delta R.T. -0.00 min
Lab File: VX007300.D
Acq: 29 Jan 2019 20:35

Tgt Ion: 105 Resp: 8833
Ion Ratio Lower Upper
105 100
120 54.3 25.1 75.3





#81

trans-1,4-Dichloro-2-butene

Concen: 62.958 ug/l

RT: 11.13 min Scan# 1637


Delta R.T. 0.06 min

Lab File: VX007300.D

Acq: 29 Jan 2019 20:35

Tot Ion: 75 Resp: 164977




Ion	Ratio	Lower	Upper
75	100		
53	0.1	79.7	119.5#
89	0.0	38.0	57.0#



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 53.00 (52.70 to 53.70): VX0

Ion 88.90 (88.60 to 89.60): VX0



#84

1,2,4-Trimethylbenzene

Concen: 1.007 ug/l

RT: 11.80 min Scan# 1747


Delta R.T. -0.00 min

Lab File: VX007300.D

Acq: 29 Jan 2019 20:35



Tgt Ion: 105 Resp: 22905


Ion	Ratio	Lower	Upper
105	100		
120	42.5	23.2	69.6



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V





#86

p-Isopropyltoluene

Concen: 0.547 ug/l





RT: 12.06 min Scan# 1789

Delta R.T. -0.00 min

Lab File: VX007300.D

Acq: 29 Jan 2019 20:35

Tot Ion:	119	Resp:	13179
Ion	Ratio	Lower	Upper
119	100		
134	31.3	14.0	41.9
91	30.1	11.3	33.9



#90

Hexachloroethane

Concen: 3.137 ug/l




RT: 12.58 min Scan# 1875


Delta R.T. -0.01 min

Lab File: VX007300.D

Acq: 29 Jan 2019 20:35


Tgt Ion:	117	Resp:	936
Ion	Ratio	Lower	Upper
117	100		
201	4.6	42.9	128.7#





#95
Naphthalene
Concen: 0.694 ug/l
RT: 13.83 min Scan# 2080
Delta R.T. -0.00 min
Lab File: VX007300.D
Acq: 29 Jan 2019 20:35

Tot Ion:128	Resp:	19185
Ion Ratio	Lower	Upper
128	100	
127	15.4	10.1 15.1#
129	11.5	8.6 13.0



Abundance Ion 128.00 (127.70 to 128.70): V
20000 Ion 127.00 (126.70 to 127.70): V
Ion 129.00 (128.70 to 129.70): V

