

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX012919\
 Data File : VX007286.D
 Acq On : 29 Jan 2019 14:22
 Operator : JC/SP
 Sample : K1262-01 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EFF-WW

Quant Time: Jan 30 09:31:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X012519W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Jan 28 10:35:22 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.02	128	77588	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.86	114	409354	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.11	117	369953	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	6.07	65	158633	31.15	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	103.83%
60) 4-Bromofluorobenzene	11.13	95	165048	28.91	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	96.37%
63) Toluene-d8	8.71	98	485483	29.11	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	97.03%

Target Compounds

					Ovalue	
15) Acetone	2.45	58	270656	372.940	ug/l	95
20) Diisopropyl ether	3.87	45	8464	0.653	ug/l #	85
25) Chloroform	5.21	83	19818	2.590	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX012919\
Data File : VX007286.D
Acq On : 29 Jan 2019 14:22
Operator : JC/SP
Sample : K1262-01 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EFF-WW

Quant Time: Jan 30 09:31:41 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X012519W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Mon Jan 28 10:35:22 2019
Response via : Initial Calibration

