

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX012919\
 Data File : VX007291.D
 Acq On : 29 Jan 2019 16:35
 Operator : JC/SP
 Sample : PB116677ZHE#02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB116677ZHE#02

Quant Time: Jan 30 03:00:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	528983	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	814892	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	792329	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	380305	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	286225	51.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.54%	
35) Dibromofluoromethane	5.50	113	255282	48.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.84%	
50) Toluene-d8	8.71	98	1000395	52.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.74%	
62) 4-Bromofluorobenzene	11.13	95	360295	54.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.20%	
Target Compounds						
16) Acetone	2.45	43	28992	10.608	ug/l	97
18) Methyl Acetate	2.78	43	14814	1.872	ug/l	94
20) Methylene Chloride	2.86	84	12104	1.783	ug/l	96
25) 2-Butanone	4.70	43	6236	1.595	ug/l	99
43) Isopropyl Acetate	6.46	43	27976	2.335	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX012919\
Data File : VX007291.D
Acq On : 29 Jan 2019 16:35
Operator : JC/SP
Sample : PB116677ZHE#02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB116677ZHE#02

Quant Time: Jan 30 03:00:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

