

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX012919\
 Data File : VX007292.D
 Acq On : 29 Jan 2019 17:02
 Operator : JC/SP
 Sample : PB116677ZHE#03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB116677ZHE#03

Quant Time: Jan 30 03:03:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	527876	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	816275	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	784279	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	370292	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	284658	51.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.20%	
35) Dibromofluoromethane	5.51	113	255166	48.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.62%	
50) Toluene-d8	8.71	98	992943	51.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.78%	
62) 4-Bromofluorobenzene	11.13	95	344569	51.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.32%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	29254	10.726	ug/l	94
18) Methyl Acetate	2.78	43	19113	2.420	ug/l	96
20) Methylene Chloride	2.86	84	12336	1.833	ug/l	97
25) 2-Butanone	4.71	43	6929	1.776	ug/l	95
43) Isopropyl Acetate	6.46	43	27110	2.259	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX012919\
Data File : VX007292.D
Acq On : 29 Jan 2019 17:02
Operator : JC/SP
Sample : PB116677ZHE#03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB116677ZHE#03

Quant Time: Jan 30 03:03:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

