

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX012919\
 Data File : VX007294.D
 Acq On : 29 Jan 2019 17:55
 Operator : JC/SP
 Sample : PB116677ZHE#05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB116677ZHE#05

Quant Time: Jan 30 03:06:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	527763	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	830911	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	796919	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	379119	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	289857	52.55	ug/l	0.00
Spiked Amount			Recovery	=	105.10%	
35) Dibromofluoromethane	5.51	113	261486	49.14	ug/l	0.00
Spiked Amount			Recovery	=	98.28%	
50) Toluene-d8	8.71	98	1009020	51.81	ug/l	0.00
Spiked Amount			Recovery	=	103.62%	
62) 4-Bromofluorobenzene	11.13	95	356491	52.50	ug/l	0.00
Spiked Amount			Recovery	=	105.00%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	33175	12.167	ug/l	97
18) Methyl Acetate	2.78	43	23747	3.007	ug/l	99
20) Methylene Chloride	2.86	84	37249	6.692	ug/l	94
25) 2-Butanone	4.70	43	7838	2.009	ug/l	92
43) Isopropyl Acetate	6.46	43	33936	2.778	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX012919\
Data File : VX007294.D
Acq On : 29 Jan 2019 17:55
Operator : JC/SP
Sample : PB116677ZHE#05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB116677ZHE#05

Quant Time: Jan 30 03:06:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

