

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX012919\
 Data File : VX007298.D
 Acq On : 29 Jan 2019 19:42
 Operator : JC/SP
 Sample : K1010-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OK-03-012819-B

Quant Time: Jan 30 03:21:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	534957	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	824233	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	805303	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	388139	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	289137	51.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.44%	
35) Dibromofluoromethane	5.51	113	261998	49.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.28%	
50) Toluene-d8	8.71	98	1016346	52.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.20%	
62) 4-Bromofluorobenzene	11.13	95	362051	53.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.50%	
Target Compounds						
					Qvalue	
16) Acetone	2.45	43	39228	14.193	ug/l	93
18) Methyl Acetate	2.78	43	20147	2.517	ug/l	92
20) Methylene Chloride	2.86	84	2452907	471.373	ug/l	94
25) 2-Butanone	4.70	43	7840	1.982	ug/l #	80
43) Isopropyl Acetate	6.46	43	31496	2.599	ug/l	99
95) Naphthalene	13.83	128	41164	1.472	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX012919\
Data File : VX007298.D
Acq On : 29 Jan 2019 19:42
Operator : JC/SP
Sample : K1010-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OK-03-012819-B

Quant Time: Jan 30 03:21:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

