

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012921\
 Data File : VX020798.D
 Acq On : 29 Jan 2021 10:40
 Operator : JC/SP
 Sample : VX0129WBL01
 Misc : 25.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK304

Quant Time: Jan 30 02:02:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXSIM012821.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Jan 30 02:01:56 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.822	114	10468	0.50	ug/L	0.00
5) Chlorobenzene-d5	10.089	117	9293	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	12.063	152	4235	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.387	65	5037	0.54	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery = 108.00%			
4) 1,2-Dichloroethane-d4	6.013	65	4272	0.47	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 94.00%			
7) 1,2-Dichloropropane-d6	7.349	67	4557	0.54	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery = 108.00%			
8) Toluene-d8	8.682	98	10414	0.45	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 90.00%			
10) 1,1,2,2-Tetrachloroeth...	11.223	84	3567	0.59	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery = 118.00%			
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.387	62	43	0.00	ug/L #	1
6) Trichloroethene	7.180	95	61	0.01	ug/L #	89
9) 1,2-Dibromoethane	9.641	107	13	0.00	ug/L #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012921\
Data File : VX020798.D
Acq On : 29 Jan 2021 10:40
Operator : JC/SP
Sample : VX0129WBL01
Misc : 25.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK304

Quant Time: Jan 30 02:02:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXSIM012821.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jan 30 02:01:56 2021
Response via : Initial Calibration

