

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012921\
 Data File : VX020800.D
 Acq On : 29 Jan 2021 12:36
 Operator : JC/SP
 Sample : VSTDCCC0.5EC
 Misc : 25.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTD0.5305

Quant Time: Jan 30 02:02:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXSIM012821.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Jan 30 02:01:56 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.822	114	9790	0.50	ug/L	0.00
5) Chlorobenzene-d5	10.092	117	8762	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	12.068	152	3831	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.387	65	5038	0.57	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	114.00%
4) 1,2-Dichloroethane-d4	6.013	65	4179	0.49	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	98.00%
7) 1,2-Dichloropropane-d6	7.368	67	4451	0.56	ug/L	0.02
Spiked Amount	0.500	Range	60 - 140	Recovery	=	112.00%
8) Toluene-d8	8.685	98	10259	0.47	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	94.00%
10) 1,1,2,2-Tetrachloroeth...	11.228	84	3251	0.57	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	114.00%
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.387	62	4920	0.58	ug/L	99
6) Trichloroethene	7.180	95	4156	0.56	ug/L	99
9) 1,2-Dibromoethane	9.644	107	2523	0.58	ug/L	99
12) 1,2,3-Trichloropropane	11.262	75	2350	0.62	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.976	75	456	0.59	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012921\
Data File : VX020800.D
Acq On : 29 Jan 2021 12:36
Operator : JC/SP
Sample : VSTDCCC0.5EC
Misc : 25.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTD0.5305

Quant Time: Jan 30 02:02:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXSIM012821.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jan 30 02:01:56 2021
Response via : Initial Calibration

