

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012924\
 Data File : VX039997.D
 Acq On : 29 Jan 2024 13:39
 Operator : JC/MD
 Sample : VX0129MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0129MBL01

Quant Time: Jan 30 02:45:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	69110	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	133307	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	132096	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	62408	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	48580	41.963	ug/l	0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	83.920%
35) Dibromofluoromethane	5.385	113	42671	47.524	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.040%
50) Toluene-d8	8.641	98	132875	39.238	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	78.480%#
62) 4-Bromofluorobenzene	11.079	95	39342	29.709	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	59.420%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012924\
Data File : VX039997.D
Acq On : 29 Jan 2024 13:39
Operator : JC/MD
Sample : VX0129MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0129MBL01

Quant Time: Jan 30 02:45:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

