

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012924\
 Data File : VX039999.D
 Acq On : 29 Jan 2024 16:21
 Operator : JC/MD
 Sample : PB158669TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB158669TB

Quant Time: Jan 30 02:46:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	72975	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	135729	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	91293	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	66893	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	59159	48.394	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	96.780%		
35) Dibromofluoromethane	5.385	113	44017	48.148	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.300%		
50) Toluene-d8	8.647	98	123370	35.781	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	71.560%#		
62) 4-Bromofluorobenzene	11.079	95	45828	33.989	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	67.980%#		
Target Compounds						
16) Acetone	2.374	43	15251	22.184	ug/l	96
20) Methylene Chloride	2.788	84	2821	2.713	ug/l	95
43) Isopropyl Acetate	6.336	43	69103	24.814	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012924\
Data File : VX039999.D
Acq On : 29 Jan 2024 16:21
Operator : JC/MD
Sample : PB158669TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB158669TB

Quant Time: Jan 30 02:46:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

