

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012924\
 Data File : VX040000.D
 Acq On : 29 Jan 2024 16:44
 Operator : JC/MD
 Sample : PB158654TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB158654TB

Quant Time: Jan 30 02:47:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	74629	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	95766	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	147637	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	67598	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	46395	37.112	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	74.220%#
35) Dibromofluoromethane	5.379	113	32568	50.491	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.980%
50) Toluene-d8	8.647	98	125052	51.404	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.800%
62) 4-Bromofluorobenzene	11.079	95	42019	44.169	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.340%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	15105	21.484	ug/l	93
20) Methylene Chloride	2.788	84	2507	2.357	ug/l #	86
43) Isopropyl Acetate	6.336	43	58843	29.947	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012924\
Data File : VX040000.D
Acq On : 29 Jan 2024 16:44
Operator : JC/MD
Sample : PB158654TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB158654TB

Quant Time: Jan 30 02:47:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

