

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012924\
 Data File : VX040002.D
 Acq On : 29 Jan 2024 17:29
 Operator : JC/MD
 Sample : P1270-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-LIQUID-20240126

Quant Time: Jan 30 02:48:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	47067	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	115483	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	82506	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	36694	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	61736	78.301	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 156.600%#			
35) Dibromofluoromethane	5.391	113	31510	40.510	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 81.020%			
50) Toluene-d8	8.647	98	113969	38.849	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 77.700%#			
62) 4-Bromofluorobenzene	11.079	95	39229	34.196	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 68.400%#			
Target Compounds					Qvalue	
16) Acetone	2.374	43	9647	21.756	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012924\
Data File : VX040002.D
Acq On : 29 Jan 2024 17:29
Operator : JC/MD
Sample : P1270-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-LIQUID-20240126

Quant Time: Jan 30 02:48:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

