

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX012924\
 Data File : VX040007.D
 Acq On : 29 Jan 2024 19:24
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 30 02:51:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	149	0.00
2 T	Dichlorodifluoromethane	50.000	19.336	61.3#	56	0.00
3 P	Chloromethane	50.000	16.267	67.5#	50	0.00
4 C	Vinyl Chloride	50.000	17.673	64.7#	54	0.00
5 T	Bromomethane	50.000	33.067	33.9#	101	0.00
6 T	Chloroethane	50.000	21.427	57.1#	62	0.00
7 T	Trichlorofluoromethane	50.000	18.870	62.3#	58	0.00
8 T	Diethyl Ether	50.000	20.065	59.9#	62	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	18.473	63.1#	56	0.00
10 T	Methyl Iodide	50.000	25.806	48.4#	73	0.00
11 T	Tert butyl alcohol	250.000	131.141	47.5#	79	0.00
12 CM	1,1-Dichloroethene	50.000	18.437	63.1#	56	0.00
13 T	Acrolein	250.000	121.362	51.5#	77	0.00
14 T	Allyl chloride	50.000	23.540	52.9#	71	0.00
15 T	Acrylonitrile	250.000	125.099	50.0#	75	0.00
16 T	Acetone	250.000	104.287	58.3#	62	0.00
17 T	Carbon Disulfide	50.000	23.307	53.4#	72	0.00
18 T	Methyl Acetate	50.000	29.651	40.7#	89	0.00
19 T	Methyl tert-butyl Ether	50.000	25.793	48.4#	79	0.00
20 T	Methylene Chloride	50.000	24.715	50.6#	77	0.00
21 T	trans-1,2-Dichloroethene	50.000	26.438	47.1#	81	0.00
22 T	Diisopropyl ether	50.000	24.782	50.4#	76	0.00
23 T	Vinyl Acetate	250.000	123.593	50.6#	73	0.00
24 P	1,1-Dichloroethane	50.000	25.533	48.9#	77	0.00
25 T	2-Butanone	250.000	120.699	51.7#	72	0.00
26 T	2,2-Dichloropropane	50.000	23.018	54.0#	70	0.00
27 T	cis-1,2-Dichloroethene	50.000	26.550	46.9#	79	0.00
28 T	Bromochloromethane	50.000	20.736	58.5#	62	0.00
29 T	Tetrahydrofuran	250.000	127.274	49.1#	76	0.00
30 C	Chloroform	50.000	27.662	44.7#	85	0.00
31 T	Cyclohexane	50.000	24.633	50.7#	75	0.00
32 T	1,1,1-Trichloroethane	50.000	27.181	45.6#	81	0.00
33 S	1,2-Dichloroethane-d4	50.000	42.301	15.4	133	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	147	0.00
35 S	Dibromofluoromethane	50.000	47.593	4.8	145	0.00
36 T	1,1-Dichloropropene	50.000	26.269	47.5#	80	0.00
37 T	Ethyl Acetate	50.000	24.887	50.2#	76	0.00
38 T	Carbon Tetrachloride	50.000	28.464	43.1#	83	0.00
39 T	Methylcyclohexane	50.000	24.533	50.9#	73	0.00
40 TM	Benzene	50.000	27.006	46.0#	81	0.00
41 T	Methacrylonitrile	50.000	27.273	45.5#	76	0.00
42 TM	1,2-Dichloroethane	50.000	27.593	44.8#	81	0.00
43 T	Isopropyl Acetate	50.000	25.650	48.7#	73	0.00
44 TM	Trichloroethene	50.000	26.101	47.8#	77	0.00
45 C	1,2-Dichloropropane	50.000	25.187	49.6#	76	0.00
46 T	Dibromomethane	50.000	26.791	46.4#	79	0.00
47 T	Bromodichloromethane	50.000	27.458	45.1#	80	0.00
48 T	Methyl methacrylate	50.000	25.375	49.3#	74	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012924\
 Data File : VX040007.D
 Acq On : 29 Jan 2024 19:24
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 30 02:51:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	520.971	47.9#	75	0.00
50 S	Toluene-d8	50.000	46.005	8.0	141	0.00
51 T	4-Methyl-2-Pentanone	250.000	129.673	48.1#	75	0.00
52 CM	Toluene	50.000	28.091	43.8#	82	0.00
53 T	t-1,3-Dichloropropene	50.000	27.648	44.7#	78	0.00
54 T	cis-1,3-Dichloropropene	50.000	26.109	47.8#	74	0.00
55 T	1,1,2-Trichloroethane	50.000	26.860	46.3#	80	0.00
56 T	Ethyl methacrylate	50.000	27.238	45.5#	78	0.00
57 T	1,3-Dichloropropane	50.000	27.182	45.6#	80	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	119.054	52.4#	68	0.00
59 T	2-Hexanone	250.000	131.635	47.3#	74	0.00
60 T	Dibromochloromethane	50.000	29.777	40.4#	85	0.00
61 T	1,2-Dibromoethane	50.000	28.677	42.6#	83	0.00
62 S	4-Bromofluorobenzene	50.000	48.409	3.2	146	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	151	0.00
64 T	Tetrachloroethene	50.000	26.260	47.5#	84	0.00
65 PM	Chlorobenzene	50.000	26.680	46.6#	82	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	28.080	43.8#	82	0.00
67 C	Ethyl Benzene	50.000	27.001	46.0#	81	0.00
68 T	m/p-Xylenes	100.000	54.931	45.1#	83	0.00
69 T	o-Xylene	50.000	27.917	44.2#	85	0.00
70 T	Styrene	50.000	28.503	43.0#	85	0.00
71 P	Bromoform	50.000	30.264	39.5#	85	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	158	0.00
73 T	Isopropylbenzene	50.000	25.511	49.0#	82	0.00
74 T	N-amyl acetate	50.000	24.702	50.6#	75	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	25.197	49.6#	81	0.00
76 T	1,2,3-Trichloropropane	50.000	24.270	51.5#	79	0.00
77 T	Bromobenzene	50.000	26.612	46.8#	87	0.00
78 T	n-propylbenzene	50.000	26.004	48.0#	82	0.00
79 T	2-Chlorotoluene	50.000	24.367	51.3#	81	0.00
80 T	1,3,5-Trimethylbenzene	50.000	25.981	48.0#	83	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	24.681	50.6#	71	0.00
82 T	4-Chlorotoluene	50.000	25.559	48.9#	83	0.00
83 T	tert-Butylbenzene	50.000	25.512	49.0#	81	0.00
84 T	1,2,4-Trimethylbenzene	50.000	26.029	47.9#	82	0.00
85 T	sec-Butylbenzene	50.000	26.057	47.9#	82	0.00
86 T	p-Isopropyltoluene	50.000	26.697	46.6#	83	0.00
87 T	1,3-Dichlorobenzene	50.000	26.501	47.0#	87	0.00
88 T	1,4-Dichlorobenzene	50.000	26.550	46.9#	88	0.00
89 T	n-Butylbenzene	50.000	26.759	46.5#	83	0.00
90 T	Hexachloroethane	50.000	26.099	47.8#	79	0.00
91 T	1,2-Dichlorobenzene	50.000	25.972	48.1#	86	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	25.569	48.9#	79	0.00
93 T	1,2,4-Trichlorobenzene	50.000	26.901	46.2#	88	0.00
94 T	Hexachlorobutadiene	50.000	23.268	53.5#	74	0.00
95 T	Naphthalene	50.000	27.507	45.0#	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012924\
Data File : VX040007.D
Acq On : 29 Jan 2024 19:24
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jan 30 02:51:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	26.722	46.6#	87	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6