

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX013019\
 Data File : VX007306.D
 Acq On : 30 Jan 2019 11:25
 Operator : JC/SP
 Sample : PB116701TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB116701TB

Quant Time: Jan 31 00:52:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	552506	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	845951	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	814955	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	391189	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	292405	50.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.28%	
35) Dibromofluoromethane	5.51	113	266849	49.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.52%	
50) Toluene-d8	8.71	98	1039148	52.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.80%	
62) 4-Bromofluorobenzene	11.13	95	367033	53.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.18%	
Target Compounds						
16) Acetone	2.45	43	35618	12.478	ug/l	96
20) Methylene Chloride	2.86	84	34700	5.892	ug/l	91
25) 2-Butanone	4.70	43	9063	2.219	ug/l	98
43) Isopropyl Acetate	6.46	43	32920	2.647	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX013019\
Data File : VX007306.D
Acq On : 30 Jan 2019 11:25
Operator : JC/SP
Sample : PB116701TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB116701TB

Quant Time: Jan 31 00:52:41 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

