

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX013019\
 Data File : VX007308.D
 Acq On : 30 Jan 2019 12:18
 Operator : JC/SP
 Sample : PB116701ZHE#02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB116701ZHE#02

Quant Time: Jan 31 02:11:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	548931	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	849876	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	820991	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	399810	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	297540	51.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.74%	
35) Dibromofluoromethane	5.50	113	267791	49.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.40%	
50) Toluene-d8	8.71	98	1039162	52.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.32%	
62) 4-Bromofluorobenzene	11.13	95	369815	53.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.50%	
Target Compounds						
16) Acetone	2.45	43	38831	13.692	ug/l	95
18) Methyl Acetate	2.78	43	18265	2.224	ug/l	98
20) Methylene Chloride	2.86	84	30326	5.114	ug/l	94
25) 2-Butanone	4.70	43	8244	2.032	ug/l #	86
43) Isopropyl Acetate	6.46	43	33312	2.666	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX013019\
Data File : VX007308.D
Acq On : 30 Jan 2019 12:18
Operator : JC/SP
Sample : PB116701ZHE#02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB116701ZHE#02

Quant Time: Jan 31 02:11:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

