

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX013019\
 Data File : VX007309.D
 Acq On : 30 Jan 2019 12:45
 Operator : JC/SP
 Sample : PB116701ZHE#03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB116701ZHE#03

Quant Time: Jan 31 03:55:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	543431	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	841374	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	807806	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	386678	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	294893	51.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.84%	
35) Dibromofluoromethane	5.51	113	266516	49.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.92%	
50) Toluene-d8	8.71	98	1025700	52.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.02%	
62) 4-Bromofluorobenzene	11.13	95	362232	52.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.36%	
Target Compounds						
16) Acetone	2.45	43	40198	14.317	ug/l	99
18) Methyl Acetate	2.78	43	25137	3.092	ug/l	97
20) Methylene Chloride	2.86	84	30492	5.203	ug/l	97
25) 2-Butanone	4.70	43	8325	2.072	ug/l #	83
43) Isopropyl Acetate	6.46	43	33736	2.727	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX013019\
Data File : VX007309.D
Acq On : 30 Jan 2019 12:45
Operator : JC/SP
Sample : PB116701ZHE#03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB116701ZHE#03

Quant Time: Jan 31 03:55:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

