

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX013019\
 Data File : VX007311.D
 Acq On : 30 Jan 2019 13:38
 Operator : JC/SP
 Sample : PB116701ZHE#05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB116701ZHE#05

Quant Time: Jan 31 03:58:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	525687	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	812472	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	789976	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	380754	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	285539	51.97	ug/l	0.00
Spiked Amount			Recovery	=	103.94%	
35) Dibromofluoromethane	5.50	113	258298	49.64	ug/l	0.00
Spiked Amount			Recovery	=	99.28%	
50) Toluene-d8	8.71	98	992388	52.11	ug/l	0.00
Spiked Amount			Recovery	=	104.22%	
62) 4-Bromofluorobenzene	11.13	95	348734	52.52	ug/l	0.00
Spiked Amount			Recovery	=	105.04%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	40571	14.938	ug/l	100
18) Methyl Acetate	2.78	43	30480	3.875	ug/l	98
20) Methylene Chloride	2.86	84	28534	5.014	ug/l	97
25) 2-Butanone	4.71	43	7304	1.879	ug/l	97
43) Isopropyl Acetate	6.45	43	32948	2.758	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX013019\
Data File : VX007311.D
Acq On : 30 Jan 2019 13:38
Operator : JC/SP
Sample : PB116701ZHE#05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB116701ZHE#05

Quant Time: Jan 31 03:58:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

