

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX013019\
 Data File : VX007312.D
 Acq On : 30 Jan 2019 14:04
 Operator : JC/SP
 Sample : PB116701ZHE#06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB116701ZHE#06

Quant Time: Jan 31 04:00:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	513078	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	800458	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	782730	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	381717	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	276801	51.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.24%	
35) Dibromofluoromethane	5.50	113	252501	49.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.52%	
50) Toluene-d8	8.71	98	988448	52.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.36%	
62) 4-Bromofluorobenzene	11.13	95	351191	53.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.38%	
Target Compounds						
16) Acetone	2.45	43	38842	14.653	ug/l	94
18) Methyl Acetate	2.78	43	28172	3.670	ug/l	94
20) Methylene Chloride	2.86	84	28356	5.116	ug/l	86
25) 2-Butanone	4.70	43	6615	1.744	ug/l #	77
43) Isopropyl Acetate	6.46	43	31313	2.661	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX013019\
Data File : VX007312.D
Acq On : 30 Jan 2019 14:04
Operator : JC/SP
Sample : PB116701ZHE#06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB116701ZHE#06

Quant Time: Jan 31 04:00:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

