

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX013019\
 Data File : VX007314.D
 Acq On : 30 Jan 2019 14:58
 Operator : JC/SP
 Sample : PB116701ZHE#08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB116701ZHE#08

Quant Time: Jan 31 04:04:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	511795	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	796212	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	780404	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	375102	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	272523	50.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.90%	
35) Dibromofluoromethane	5.51	113	245990	48.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.48%	
50) Toluene-d8	8.71	98	980274	52.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.04%	
62) 4-Bromofluorobenzene	11.13	95	348724	53.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.20%	
Target Compounds						
16) Acetone	2.45	43	35914	13.582	ug/l	97
18) Methyl Acetate	2.78	43	28768	3.757	ug/l	95
20) Methylene Chloride	2.86	84	33344	6.133	ug/l	95
25) 2-Butanone	4.71	43	7774	2.055	ug/l #	87
43) Isopropyl Acetate	6.46	43	31774	2.714	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX013019\
Data File : VX007314.D
Acq On : 30 Jan 2019 14:58
Operator : JC/SP
Sample : PB116701ZHE#08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB116701ZHE#08

Quant Time: Jan 31 04:04:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

