

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX013019\
 Data File : VX007316.D
 Acq On : 30 Jan 2019 15:51
 Operator : JC/SP
 Sample : K1282-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-2

Quant Time: Jan 31 04:09:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	517818	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	800068	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	779920	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	369162	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	279988	51.74	ug/l	0.00
Spiked Amount			Recovery	=	103.48%	
35) Dibromofluoromethane	5.50	113	252677	49.32	ug/l	0.00
Spiked Amount			Recovery	=	98.64%	
50) Toluene-d8	8.71	98	983847	52.46	ug/l	0.00
Spiked Amount			Recovery	=	104.92%	
62) 4-Bromofluorobenzene	11.13	95	342672	52.41	ug/l	0.00
Spiked Amount			Recovery	=	104.82%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	52233	19.524	ug/l	99
18) Methyl Acetate	2.78	43	25901	3.343	ug/l	95
20) Methylene Chloride	2.86	84	458504	90.565	ug/l	93
25) 2-Butanone	4.70	43	18486	4.829	ug/l	96
43) Isopropyl Acetate	6.46	43	26534	2.256	ug/l	95
95) Naphthalene	13.83	128	73045	2.746	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX013019\
Data File : VX007316.D
Acq On : 30 Jan 2019 15:51
Operator : JC/SP
Sample : K1282-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
TP-2

Quant Time: Jan 31 04:09:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

