

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX013020\
 Data File : VX014772.D
 Acq On : 30 Jan 2020 17:21
 Operator : JC/SP
 Sample : L1332-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KIT-SINK-1ST-DRAW

Quant Time: Jan 31 05:26:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 22 12:35:15 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	437302	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	6.84	114	688145	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	624651	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	268996	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	262596	51.42	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.84%	
35) Dibromofluoromethane	5.48	113	217206	50.02	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.04%	
50) Toluene-d8	8.71	98	814432	48.79	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.58%	
62) 4-Bromofluorobenzene	11.13	95	270901	45.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.12%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX013020\
Data File : VX014772.D
Acq On : 30 Jan 2020 17:21
Operator : JC/SP
Sample : L1332-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KIT-SINK-1ST-DRAW

Quant Time: Jan 31 05:26:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X012220W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 22 12:35:15 2020
Response via : Initial Calibration

