

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX013020\
 Data File : VX014768.D
 Acq On : 30 Jan 2020 10:17
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 31 04:29:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 22 12:35:15 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	100	0.00
2 T	Dichlorodifluoromethane	0.598	0.627	-4.8	90	0.00
3 P	Chloromethane	0.676	0.613	9.3	87	0.00
4 C	Vinyl Chloride	0.731	0.709	3.0#	88	0.00
5 T	Bromomethane	0.497	0.437	12.1	88	0.00
6 T	Chloroethane	0.458	0.402	12.2	90	0.00
7 T	Trichlorofluoromethane	0.852	0.809	5.0	95	0.00
8 T	Diethyl Ether	0.372	0.347	6.7	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.500	0.489	2.2	97	0.00
10 T	Methyl Iodide	0.656	0.651	0.8	89	0.00
11 T	Tert butyl alcohol	0.131	0.099	24.4#	86	0.00
12 CM	1,1-Dichloroethene	0.497	0.463	6.8#	93	0.00
13 T	Acrolein	0.103	0.108	-4.9	106	0.00
14 T	Allyl chloride	0.893	0.834	6.6	90	0.00
15 T	Acrylonitrile	0.303	0.273	9.9	90	0.00
16 T	Acetone	0.348	0.306	12.1	96	0.00
17 T	Carbon Disulfide	1.358	1.209	11.0	87	0.00
18 T	Methyl Acetate	0.913	0.830	9.1	90	0.00
19 T	Methyl tert-butyl Ether	1.628	1.587	2.5	95	0.00
20 T	Methylene Chloride	0.592	0.540	8.8	94	0.00
21 T	trans-1,2-Dichloroethene	0.547	0.503	8.0	93	0.00
22 T	Diisopropyl ether	1.777	1.754	1.3	94	0.00
23 T	Vinyl Acetate	1.355	1.344	0.8	97	0.00
24 P	1,1-Dichloroethane	0.965	0.930	3.6	94	0.00
25 T	2-Butanone	0.458	0.419	8.5	92	0.00
26 T	2,2-Dichloropropane	0.773	0.756	2.2	96	0.00
27 T	cis-1,2-Dichloroethene	0.613	0.583	4.9	95	0.00
28 T	Bromochloromethane	0.424	0.379	10.6	94	0.00
29 T	Tetrahydrofuran	0.272	0.240	11.8	87	0.00
30 C	Chloroform	0.956	0.924	3.3#	95	0.00
31 T	Cyclohexane	0.908	0.845	6.9	89	0.00
32 T	1,1,1-Trichloroethane	0.808	0.781	3.3	96	0.00
33 S	1,2-Dichloroethane-d4	0.584	0.565	3.3	98	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	96	0.00
35 S	Dibromofluoromethane	0.315	0.320	-1.6	98	0.00
36 T	1,1-Dichloropropene	0.465	0.473	-1.7	92	0.00
37 T	Ethyl Acetate	0.532	0.500	6.0	89	0.00
38 T	Carbon Tetrachloride	0.444	0.462	-4.1	96	0.00
39 T	Methylcyclohexane	0.582	0.598	-2.7	93	0.00
40 TM	Benzene	1.436	1.449	-0.9	93	0.00
41 T	Methacrylonitrile	0.319	0.281	11.9	88	0.00
42 TM	1,2-Dichloroethane	0.496	0.516	-4.0	97	0.00
43 T	Isopropyl Acetate	0.871	0.842	3.3	91	0.00
44 TM	Trichloroethene	0.405	0.405	0.0	94	0.00
45 C	1,2-Dichloropropane	0.370	0.384	-3.8#	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX013020\
 Data File : VX014768.D
 Acq On : 30 Jan 2020 10:17
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 31 04:29:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 22 12:35:15 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.240	0.247	-2.9	95	0.00
47 T	Bromodichloromethane	0.470	0.500	-6.4	97	0.00
48 T	Methyl methacrylate	0.416	0.422	-1.4	90	0.00
49 T	1,4-Dioxane	0.009	0.008	11.1	87	0.00
50 S	Toluene-d8	1.213	1.219	-0.5	96	0.00
51 T	4-Methyl-2-Pentanone	0.550	0.527	4.2	90	0.00
52 CM	Toluene	0.882	0.910	-3.2#	94	0.00
53 T	t-1,3-Dichloropropene	0.506	0.545	-7.7	96	0.00
54 T	cis-1,3-Dichloropropene	0.556	0.608	-9.4	96	0.00
55 T	1,1,2-Trichloroethane	0.361	0.372	-3.0	94	0.00
56 T	Ethyl methacrylate	0.544	0.581	-6.8	93	0.00
57 T	1,3-Dichloropropane	0.618	0.636	-2.9	96	0.00
58 T	2-Chloroethyl Vinyl ether	0.182	0.195	-7.1	93	0.00
59 T	2-Hexanone	0.427	0.415	2.8	90	0.00
60 T	Dibromochloromethane	0.377	0.398	-5.6	96	0.00
61 T	1,2-Dibromoethane	0.382	0.388	-1.6	95	0.00
62 S	4-Bromofluorobenzene	0.432	0.438	-1.4	99	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	96	0.00
64 T	Tetrachloroethene	0.495	0.491	0.8	90	0.00
65 PM	Chlorobenzene	1.069	1.083	-1.3	96	0.00
66 T	1,1,1,2-Tetrachloroethane	0.394	0.409	-3.8	98	0.00
67 C	Ethyl Benzene	1.826	1.915	-4.9#	95	0.00
68 T	m/p-Xylenes	0.692	0.748	-8.1	97	0.00
69 T	o-Xylene	0.668	0.701	-4.9	95	0.00
70 T	Styrene	1.120	1.235	-10.3	96	0.00
71 P	Bromoform	0.314	0.335	-6.7	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
73 T	Isopropylbenzene	3.516	3.610	-2.7	97	0.00
74 T	N-amyl acetate	1.541	1.523	1.2	92	0.00
75 P	1,1,2,2-Tetrachloroethane	1.255	1.145	8.8	95	0.00
76 T	1,2,3-Trichloropropane	1.163	0.968	16.8	81	0.00
77 T	Bromobenzene	0.944	0.943	0.1	98	0.00
78 T	n-propylbenzene	3.991	4.159	-4.2	97	0.00
79 T	2-Chlorotoluene	2.415	2.423	-0.3	96	0.00
80 T	1,3,5-Trimethylbenzene	2.918	3.053	-4.6	97	0.00
81 T	trans-1,4-Dichloro-2-butene	0.409	0.387	5.4	91	0.00
82 T	4-Chlorotoluene	2.749	2.839	-3.3	98	0.00
83 T	tert-Butylbenzene	2.742	2.848	-3.9	101	0.00
84 T	1,2,4-Trimethylbenzene	2.927	3.065	-4.7	98	0.00
85 T	sec-Butylbenzene	3.359	3.545	-5.5	98	0.00
86 T	p-Isopropyltoluene	3.104	3.314	-6.8	99	0.00
87 T	1,3-Dichlorobenzene	1.697	1.679	1.1	98	0.00
88 T	1,4-Dichlorobenzene	1.700	1.688	0.7	99	0.00
89 T	n-Butylbenzene	2.704	2.892	-7.0	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX013020\
Data File : VX014768.D
Acq On : 30 Jan 2020 10:17
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jan 31 04:29:58 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012220W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 22 12:35:15 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.577	0.575	0.3	99	0.00
91 T	1,2-Dichlorobenzene	1.686	1.679	0.4	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.270	0.237	12.2	90	0.00
93 T	1,2,4-Trichlorobenzene	1.087	1.154	-6.2	100	0.00
94 T	Hexachlorobutadiene	0.559	0.588	-5.2	101	0.00
95 T	Naphthalene	3.127	3.293	-5.3	94	0.00
96 T	1,2,3-Trichlorobenzene	1.099	1.160	-5.6	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6